



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor VII Deficient Plasma
Product Code : DP030A/K
Lot No. : FC02601H
Expiration date : 2025-07-21

Components	Quantity	Lot	Expiration Date
Factor VII Deficient Plasma	1 vial (DP030A) 6 vials (DP030K)	FC0260JR1H	2025-07-21

Analytical Data	Specification	Result
Calibration curve	CT (0%) >100 SEC CT (100%) <40 SEC Thromboplastin: STAGO	139sec 28sec
Concentration FVII	< 1%	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

23 FEV. 2023

S. LECOURT