



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description** EASYPLASMA CALIBRATOR

**Product Code :** 226601

**Lot No. :** FC09391B

**Expiration date :** 2025-09-09

Components	Quantity	Lot	Expiration Date
Plasma Calibrator	6 vials	FC0939bj1B	2025-09-09

Analytical Data	Specification	Result
Fibrinogen	2 - 4 g/L	3.00 g/L
Thrombin time Ratio	≤ 1.25	PASS
PT %	> 60 %	PASS
aPTT LR Ratio	≤ 1.25	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2023-04-20

Quality Control Manager : S.LECOURT