



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

<b>Product Description:</b>	<b>LYOPHILIZED PLATELET</b>
<b>Product Code:</b>	<b>AG006K</b>
<b>Lot No.:</b>	<b>FC1620</b>
<b>Expiration date:</b>	<b>2025-12-26</b>

Components	Quantity	Lot	Expiration Date
LYOPHILIZED PLATELET	3 vials	435-44	2026-03-31
TBS Buffer	3 vials	FC1620	2025-12-26

Analytical Data	Specification	Result
Solubility	≤ 10 minutes	PASS
Ref curve linearity	≥ 0.95 r <sup>2</sup>	0.98r <sup>2</sup>
100% ref point	≥ 50%/min	77%/min
pH TBS Buffer	7.25-7.45	7.33

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2023-08-02

Quality Control Manager: M.PEYRAFITTE

