



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

CERTIFICATE OF ANALYSIS	
<b>Product Description :</b>	<b>PT Calibrator Set</b>
<b>Product Code :</b>	<b>SC084K</b>
<b>Lot No. :</b>	<b>FD058218</b>
<b>Expiration date :</b>	<b>2025-12-30</b>

Components	Quantity	Lot	Expiration Date
CAL 1	2 vials	FD0582di18	2025-12-30
CAL 2	2 vials	FD0582dj18	2025-12-30
CAL 3	2 vials	FD0582dk18	2025-12-30
CAL 4	2 vials	FD0582dl18	2025-12-30
CAL 5	2 vials	FD0582dm18	2025-12-30

Analytical Data	Specification	Result
CAL 1	INR : 0.9-1.1	PASS*
CAL 2	INR : 1.8-2.2	PASS*
CAL 3	INR : 2.6-3.4	PASS*
CAL 4	INR : 3.4-4.6	PASS*
CAL 5	INR : 4.3-5.7	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

\* NCI 2365 : Accepted under waiver

Date of Approval : 2024-06-20

Quality Control Manager :

Sophie LECOURT  
Deputy Director Quality Control

