



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ZYMUTEST PAI-1 Activity

**Product Code :** RK019A

**Lot No. :** FC1067

**Expiration date :** 2026-01-31

Components	Quantity	Lot	Expiration Date
tPA pre-coated plate	12x8 wells	FC1250	2026-10-20
(h) PAI-1 standard	3 vials	FC1171	2026-10-08
(h) PAI-1 Control I	1 vial	FC1173	2026-10-08
(h) PAI-1 Control II	1 vial	FC1174	2026-10-08
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	FC1278	2026-10-27
F-Sample diluent	2x50 ml	FC1249	2026-10-20
Wash solution	1x50 ml	FC0035	2026-06-15
Conjugate diluent	1x25 ml	FB1457	2026-02-26
TMB	1x25 ml	220106D3	2026-01-31
Sulfuric Acid 0,45M	1x6 ml	FC0222	2026-07-15

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml $\geq$ 1.00	1.23
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.003
Detection threshold	$\leq$ 0.5 ng/ml	< 0.5 ng/mL

### Calibration Curve for information

						Acceptance Range	MV	
Conc.	C	C/2	C/4	C/10	C/20	CI	[ 7.9 - 10.7 ]	9.9
OD	1.44	0.90	0.50	0.21	0.11	CII	[ 2.6 - 4.01 ]	3.1

MV = Measured value

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2023-06-01

Quality Control Manager :

**S.LECOURT**

