



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ARACHIDONIC ACID
Product Code : AG003K
Lot No. : FD1589M6 / FD1593
Expiration date : 2026-02-13

Components	Quantity	Lot	Expiration Date
ARACHIDONIC ACID	3 vials	FD1589M6 / FD1593	2026-02-13

Analytical Data	Specification	Result
Reactivity (% AgregMax) 1.0 mM on CS or CN series	≥ 50%	82%
Reactivity (% AgregMax) 1.0 mM on Agregometer	≥ 50%	90%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

23 SEP. 2024
S. LECOURT

Quality Control Manager :

