



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PT-Phen™ LRT
Product Code : CK584K/L
Lot No. : FD1234KZ
Expiration date : 2026-02-21

Components	Quantity	Lot	Expiration Date
R1 : PT-Phen™ LRT	6 vials (CK584K) 12 vials (CK584L)	FD0255doHW	2026-02-21

Analytical Data	Specification	Result
PT on Normal Plasma	11 ± 1 sec	11.4 sec
ISI (WHO Method)	1.0 ± 0.1	0.97 on CS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

20 JUIN 2024

S. LECOURT