



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: ZYMUTEST Anti Annexin V IgG
Product Code: RK005A
Lot No.: FC1987
Expiration date: 2026-02-20

Components	Quantity	Lot	Expiration Date
Annexin V pre-coated plate	12x8 wells	FC1929	2026-02-20
Anti Annexin V IgG Standard	3 vials	FC1930	2026-02-23
Negative Control	3 vials	FC1414	2026-11-17
Anti (h) IgG Gamma HRP immunoconjugate	3 vials	FC1988	2027-02-22
Autoimmunity Sample diluent	2x50 ml	FC1412	2026-11-16
Wash solution	1x50 ml	FC1080	2026-09-22
Conjugate diluent	1x25 ml	FC1584	2026-12-16
TMB	1x25 ml	221201D01	2026-12-31
Sulfuric Acid 0,45M	1x6 ml	FC1736	2027-01-04

Analytical Data	Specification	Result
Reactivity	A450 for 50 AU/ml \geq 1.50	1.76
Calibration Curve and Negative Control	Negative Control measured within acceptance range	PASS
Blank	< 0.100	0.009
Normal Plasma	\leq 10 AU/mL	< 10 AU/mL
Detection threshold	\leq 5 AU/mL	< 1 AU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2023-09-29

Quality Control Manage: S. LECOURT

