



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : CEPHEN 1 LS
Product Code : CK521K
Lot No. : FD0514JE
Expiration date : 2026-03-10

Components	Quantity	Lot	Expiration Date
CEPHEN 1 LS	6 vials	FD0514KDJE	2026-03-10

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	1.04 on STAR 1.05 on CS 1.04 on CN

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

11 AVR. 2024
S. LECOURT

Quality Control Manager :