



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** CEPHEN 2.5  
**Product Code :** CK512K  
**Lot No. :** FD0685JN  
**Expiration date :** 2026-04-06

Components	Quantity	Lot	Expiration Date
CEPHEN 2.5	6 vials	FD0685KBJN	2026-04-06

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	0.98 on STAR 0.97 on CS 0.98 on CN

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval** Quality Control Manager :

**02 MAI 2024**

**S. LECOURT**