



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL PLASMA
Product Code : 223001
Lot No. : FD1070
Expiration date : 2026-05-23

Components	Quantity	Lot	Expiration Date
C3	6 vials	FC2440BOHF	2026-05-23
C4	6 vials	FD0722BPJU	2026-09-26

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml C4 : 1.05 -1.35 IU/ml	0.78 IU/mL 1.16 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2024-05-24

Quality Control Manager : po SDE

M. Peyrache
Tech. Supp. spc.

