

Certificate of Analysis

Product: 5-PLASMA Normal Plasma 5

Reference Number: 5D-40401B

Lot Number: 418812

Expiration Date: 2026-05-25

Storage: 2-8°C


Presentation: 8 vials of 5 mL

Specifications:

Analytical Data	Specifications	Result
INR	< 1,3	1.18
aPTT ratio	< 1.20	1.07
Fibrinogen	≥ 2 g/L	2.41 g/L
FVIII:C	> 50%	82%
LA DVV test / DVV confirm ratio	≤ 1.20	Pass

For Research Use Only. Not for use in diagnostic procedures. For *in vitro* use only.

Signature:



Urs-Peter Geiger
Date: 2024-04-29