



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : EPINEPHRINE
Product Code : AG002K
Lot No. : FC2331
Expiration date : 2026-05-05

Components	Quantity	Lot	Expiration Date
EPINEPHRINE	3 vials	FC2331	2026-05-05

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 5 μ M on CS or CN series	$\geq 70\%$	89%
Reactivity (% AgregMax) at 5 μ M on Agregometer	$\geq 70\%$	78%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-11-27

Quality Control Manager : S. LECOURT

