



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS
Product Description : ZYMUTEST PF4
Product Code : RK006A
Lot No. : FD0050
Expiration date : 2026-06-26

Components	Quantity	Lot	Expiration Date
PF4 pre-coated plate	12x8 wells	FD0324	2026-07-31
PF4 Standard	3 vials	FD0042	2026-06-26
PF4 Control I	1 vial	FD0041	2026-06-26
PF4 Control II	1 vial	FD0045	2026-06-26
PF4 HRP immunoconjugate	3 vials	FD0326	2026-08-03
PF4 Sample diluent	2x50 ml	FD0408	2026-08-07
Wash solution	1x50 ml	FD0322	2027-07-26
Conjugate diluent	1x25 ml	FC1584	2026-12-16
TMB	1x25 ml	221201D01	2026-12-31
Sulfuric Acid 0,45M	1x6 ml	FC2199	2027-03-30

Analytical Data	Specification	Result
Reactivity	A450 for 8 IU/ml \geq 1.70	1,8
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0,093
Detection threshold	\leq 0.5 IU/ml	< 0,043

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

04 MARS 2024
S. LECOURT
