



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description : BIOPHEN RIVAROXABAN CALIBRATOR**

**Product Code : 222701**

**Lot No. : FD00981E / FD009715**

**Expiration date : 2026-07-04**

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FD0098BA1E / FD009715	2026-07-04
CAL 2	4 vials	FD0098BB1E / FD009715	2026-07-04
CAL 3	4 vials	FD0098BC1E / FD009715	2026-07-04

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -500 ng/ml	0 250 460
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Quality Control Approval

Date of Approval :

05 FEV. 2024

Quality Control Manager :

S. LECOURT

