



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	BIOPHEN™ Rivaroxaban Control Low
Product Code :	225101
Lot No. :	FD02021G
Expiration date :	2026-07-18

Components	Quantity	Lot	Expiration Date
CI	6 vials	FD0202DM1G	2026-07-18
CII	6 vials	FD0202DN1G	2026-07-18

Analytical Data	Specification	Result
Concentration	CI : 15 -35 ng/ml CII : 45 -105 ng/ml	25 ng/ml 61 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

26 FEV. 2024

S. LECOURT

Quality Control Manager :