



 HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN APIXABAN CONTROL LOW  
**Product Code :** 225201  
**Lot No. :** FD02631G  
**Expiration date :** 2026-07-27

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD0263DW1G	2026-07-27
C2	6 vials	FD0263DX1G	2026-07-27

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 45 -105 ng/ml	24 ng/ml 63 ng/ml

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

**01 MARS 2024**

Quality Control Manager :

  
Sophie LECOURT  
Deputy Director Quality Control