



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CALIBRATOR

Product Code : 226201

Lot No. : FD03071E

Expiration date : 2026-08-04

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FD0307DO1E	2026-08-04
CAL 2	4 vials	FD0307DP1E	2026-08-04
CAL 3	4 vials	FD0307DQ1E	2026-08-04

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -400 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/ml 282 ng/ml 535 ng/ml
Linearity and accuracy of calibration with values assigned	R ² ≥ 0.98 Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

04 AVR. 2024

S. LECOURT

