



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ADP
Product Code : AG001K
Lot No. : FD0442J8 / FD0444
Expiration date : 2026-08-15

Components	Quantity	Lot	Expiration Date
ADP	3 vials	FD0442J8 / FD0444	2026-08-15

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 2µM on CS or CN series	≥ 70%	87%
Reactivity (% AgregMax) at 2µM on agregometer	≥ 70%	82%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

19 AVR. 2024

S. LECOURT

Quality Control Manager :

