



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** RISTOCETIN  
**Product Code :** AG004K  
**Lot No. :** FD0394J5 / FD0395  
**Expiration date :** 2026-08-08

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	FD0394J5 / FD0395	2026-08-08

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.2mg/ml on CS or CN series	≥ 50%	80%
Reactivity (% AgregMax) at 1.2mg/ml on Agregometer	≥ 50%	85%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval** Quality Control Manager :

20 MARS 2024

S. LECOURT

