



 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : COLLAGEN
Product Code : AG005K
Lot No. : FD0345J2 / FD0348
Expiration date : 2026-08-01

Components	Quantity	Lot	Expiration Date
COLLAGEN REAGENT	3 vials	FD0345J2 / FD0348	2026-08-01
COLLAGEN DILUENT	3 vials	FD0345J2 / FD0348	2026-08-14

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 2 µg/ml on CS or CN Series	≥ 70%	81%
Reactivity (% AgregMax) at 2 µg/ml on Agregometer	≥ 70%	79%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

25 MARS 2024

S. LECOURT

Quality Control Manager :