



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

<b>Product Description :</b>	<b>IMIDAZOLE BUFFER</b>
<b>Product Code :</b>	<b>AR021K/L</b>
<b>Lot No. :</b>	<b>FD0387J4</b>
<b>Expiration date :</b>	<b>2026-08-07</b>

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	3 vials *25mL (AR021K) 12 vials *25mL (AR021L)	FD0387ARJ4	2026-08-07

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample $\leq$ 5%	<1%
FVIII:C clotting assay	Delta CT (100%) with control sample $\leq$ 5%	<1%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**

**27 MARS 2024**  
**S. LECOURT**

Quality Control Manager :

