



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : vWF Deficient Plasma
Product Code : DP150A/K
Lot No. : FD051112
Expiration date : 2026-08-24

Components	Quantity	Lot	Expiration Date
vWF Deficient Plasma	1 vial (DP150A) 6 vials (DP150K)	FD051112	2026-08-24

Analytical Data	Specification	Result
Concentration vWF	< 1 %	<1%
Concentration FVIII	< 10%	<1%
Concentration FV	> 50 %	71%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

20 MARS 2024
S. LECOURT

