



HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN PROTEIN C LRT
Product Code : 221211
Lot No. : FD07651M
Expiration date : 2026-09-25

Components	Quantity	Lot	Expiration Date
R1 : Protac	3 vials	FD0765a61M	2026-09-25
R2 : PC substrate	3 vials	FD0765a71M	2026-09-25

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	$A_{405}(100\%) = 0.12 \pm 0.04$ $\Delta A_{405}(100-50\%) > 0.03$ $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.120 0.058 0.998 PASS
Detection Threshold	$\leq 5\%$	$< 1\%$

Calibration Curve for information

CS						Acceptance Range	MV	
Conc.	150	100	50	0	R^2	NC	[80-100]	91%
OD	0.173	0.120	0.062	0	0.998	AC	[30-40]	36%

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

02 MAI 2024

Quality Control Manager :

S. LECOURT