



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	BIOPHEN UFH CONTROL C2
Product Code :	223901
Lot No. :	FD0708JT
Expiration date :	2026-09-22

Components	Quantity	Lot	Expiration Date
C2	12 vials	FD0708BYJT	2026-09-22

Analytical Data	Specification	Result
Concentration	C2 : 0.40 -0.60 IU/mL	0.54 IU/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval
Date of Approval : **19 AVR. 2024**
S. LECOURT
Quality Control Manager : 