



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL LOW CI
Product Code : 224301
Lot No. : FD0647JL
Expiration date : 2026-09-15

Components	Quantity	Lot	Expiration Date
CI	12 vials	FD0647BQJL	2026-09-15

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml	0.26 IU/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

19 AVR. 2024
S. LECOURT

Quality Control Manager :