



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS
Product Description : ZYMUTEST TOTAL TAFI-Ag
Product Code : RK008A
Lot No. : FD0673
Expiration date : 2026-09-21

Components	Quantity	Lot	Expiration Date
Anti-(h) Total TAFI pre-coated plate	12x8 wells	FD0727	2026-09-25
TAFI standard	3 vials	FD0674	2026-09-21
Control I	1 vial	FD0675	2026-09-21
Control II	1 vial	FD0676	2026-09-21
Anti-(h)- Total TAFI HRP immunoconjugate	3 vials	FD0900	2026-10-19
Fibrinolysis Sample diluent	2x50 ml	FD0480	2027-08-16
Wash solution	1x50 ml	FD0073	2027-06-22
Conjugate diluent	1x25 ml	FC1584	2026-12-16
TMB	1x25 ml	221201D01	2026-12-31
Sulfuric Acid 0,45M	1x6 ml	FC2199	2027-03-30

Analytical Data	Specification	Result
Reactivity	A450 for 100% \geq 1.35	1.66
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.027
Detection threshold	\leq 5 %	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

Quality Control Manager :

16 MAI 2024
S. LECOURT
