



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST Fibrinogen

Product Code : RK024A

Lot No. : FD0608

Expiration date : 2026-09-13

Components	Quantity	Lot	Expiration Date
pre-coated plate	12x8 wells	FD0687	2026-09-19
Fibrinogen Standard	3 vials	FD0635	2026-09-13
Fibrinogen Control I	1 vial	FD0637	2026-09-13
Fibrinogen Control II	1 vial	FD0639	2026-09-13
Fibrinogen HRP immunoconjugate	3 vials	FD0811	2026-10-05
B2F Sample diluent	2x50 ml	FD0689	2026-09-19
Wash solution	1x50 ml	FD0322	2027-07-26
B2F Conjugate diluent	1x25 ml	FD0691	2026-09-19
TMB	1x25 ml	221201D01	2026-12-31
Sulfuric Acid 0,45M	1x6 ml	FD0462	2027-08-09

Analytical Data	Specification	Result
Reactivity	A450 for 40 ng/ml \geq 1.45	1.82
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.150	0.017
Detection threshold	\leq 1 ng/ml	\leq 1 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

07 MAI 2024

S. LECOURT

