



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** PLASMA ARGATROBAN  
CALIBRATOR  
**Product Code :** SC030K  
**Lot No. :** FD057519  
**Expiration date :** 2026-09-05

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FD0575BZ19	2026-09-05
CAL 2	4 vials	FD0575CA19	2026-09-05
CAL 3	4 vials	FD0575CB19	2026-09-05
CAL 4	4 vials	FD0575CC19	2026-09-05
CAL 5	4 vials	FD0575CD19	2026-09-05

Analytical Data	Specification	Result
CAL 1	$\leq 0.05 \mu\text{g/mL}$	0.00 $\mu\text{g/mL}$
CAL 2	0.35 - 0.65 $\mu\text{g/mL}$	0.52 $\mu\text{g/mL}$
CAL 3	0.85 - 1.15 $\mu\text{g/mL}$	0.97 $\mu\text{g/mL}$
CAL 4	1.30 - 1.65 $\mu\text{g/mL}$	1.49 $\mu\text{g/mL}$
CAL 5	1.70 - 2.00 $\mu\text{g/mL}$	1.93 $\mu\text{g/mL}$

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**

09 AVR. 2024  
S. LECOURT

Quality Control Manager :