



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	BIOPHEN ARIXTRA CALIBRATOR
Product Code :	222501
Lot No. :	FD090219
Expiration date :	2026-10-20

Components	Quantity	Lot	Expiration Date
CAL 1	3 vials	FD0902CN19	2026-10-20
CAL 2	3 vials	FD0902CO19	2026-10-20
CAL 3	3 vials	FD0902CP19	2026-10-20
CAL 4	3 vials	FD0902CQ19	2026-10-20

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 0.05 \mu\text{g/ml}$ CAL 2 : $0.40 - 0.60 \mu\text{g/ml}$ CAL 3 : $0.90 - 1.10 \mu\text{g/ml}$ CAL 4 : $1.35 - 1.60 \mu\text{g/ml}$	0.00 $\mu\text{g/ml}$ 0.51 $\mu\text{g/ml}$ 1.01 $\mu\text{g/ml}$ 1.45 $\mu\text{g/ml}$
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**
22 MAI 2024
S. LECOURT

Quality Control Manager : 