



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CONTROL PLASMA
Product Code : 224501
Lot No. : FD09391F
Expiration date : 2026-10-31

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD0939BD1F	2026-10-31
C2	6 vials	FD0939BE1F	2026-10-31

Analytical Data	Specification	Result
Concentration	C1 : 50 -150 ng/ml C2 : 250 -350 ng/ml	96 ng/mL 311ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval Quality Control Manager :

04 JUIN 2024

S. LECOURT