



 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ Dabigatran Control Plasma
Product Code : 224701
Lot No. : FD08021F
Expiration date : 2026-10-03

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD0802AV1F	2026-10-03
C2	6 vials	FD0802AW1F	2026-10-03

Analytical Data	Specification	Result
Concentration	C1 : 50 -200 ng/ml C2 : 200 -300 ng/ml	102 ng/ml 251 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

02 MAI 2024
S. LECOURT

Quality Control Manager :

