



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CALIBRATOR
Product Code : 226501
Lot No. : FD08591E
Expiration date : 2026-10-13

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FD0859aa1E	2026-10-13
CAL 2	4 vials	FD0859ab1E	2026-10-13
CAL 3	4 vials	FD0859ac1E	2026-10-13

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 40 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/ml 238 ng/ml 487 ng/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval : **16 MAI 2024** Quality Control Manager :

S. LECOURT

