



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ZYMUPHEN MP-TF

**Product Code :** 521196

**Lot No. :** FD0870

**Expiration date :** 2026-10-09

Components	Quantity	Lot	Expiration Date
Anti(h) TF pre-coated plate	12x8 wells	FD0944	2026-10-30
MP-TF Calibrator	2 vials	FD1151	2026-10-09
R1	2 vials	FD0980	2026-11-02
R2	2 vials	FD1037	2026-11-16
R3	2 vials	FD0943	2026-11-02
Sample Diluent SD MP-TF	1x50 ml	FD0847	2026-10-09
MP Activity Control High	1 vial	FD1152	2026-11-23
MP Activity Control Low	1 vial	FD1154	2026-11-23
Assay Enhancer AE MP-TF	1x25 ml	FD0849	2026-10-09
Wash Solution 10X WS MP-TF	1x50 ml	FD0851	2026-10-09
Citric Acid 2%	1x6 ml	FD0845	2027-10-04

Analytical Data	Specification	Result
Reactivity	A405 at 22 pg/ml $\geq$ 1.30	1.34
Calibration Curve and Controls	Controls measured within acceptance range	PASS
Blank	$\leq$ 0.200	0.16
Detection Limit	$\leq$ 1 pg/mL	< 1 pg/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

Quality Control Manager :

**Quality Control Approval**

08 JUL. 2024

S. LECOURT