



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description BIOPHEN ABNORMAL CONTROL PLASMA

Product Code : 223301

Lot No. : FD2323PB

Expiration date : 2026-11-29

Components	Quantity	Lot	Expiration Date
Abnormal Control	12 vials	FD1146EAKP	2026-11-29

Analytical Data	Specification	Result
INR	> 1,3	PASS
aPTT LR Ratio	> 1.50	PASS
Fibrinogen (Activity)	0.8-1.2 g/L	PASS
Fibrinogen (Antigen)	0.7-1.3 g/L	PASS
LA ratio	≤ 1.2	PASS
Alpha 2 Anti plasmin	20 - 50 %	PASS
Antithrombin (Activity)	20 - 50 %	PASS
Antithrombin (Antigen)	20 - 50 %	PASS
Factor II	20 - 50 %	PASS
Factor V	20 - 50 %	PASS
Factor VII	20 - 50 %	PASS
Factor X	20 - 50 %	PASS
Factor VII+ X	20 - 50 %	PASS
Factor VIII	20 - 50 %	PASS
Factor IX	20 - 50 %	PASS
Factor XI	20 - 50 %	PASS
Factor XII	20 - 50 %	PASS
Factor XIII	20 - 50 %	PASS
Free Protein S (Antigen)	20 - 50 %	PASS
Plasminogen	20 - 50 %	PASS
Protein C Chromogenic	20 - 50 %	PASS
Protein C (Antigen)	20 - 50 %	PASS
Protein C Clotting	20 - 50 %	PASS
Protein S (Activity)	20 - 50 %	PASS
von Willebrand Factor (Antigen)	20 - 50 %	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

19 DEC. 2024

S. LECOURT