

HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS


| | |
|------------------------------|-------------------------|
| Product Description : | IMIDAZOLE BUFFER |
| Product Code : | AR021K/L |
| Lot No. : | FD1155KR |
| Expiration date : | 2026-11-27 |

| Components | Quantity | Lot | Expiration Date |
|------------------|---|------------|-----------------|
| IMIDAZOLE BUFFER | 3 vials *25mL (AR021K) 12 vials *25mL (AR021L) | FD1155ARKR | 2026-11-27 |

| Analytical Data | Specification | Result |
|----------------------------|---|--------|
| Hemoclot PC clotting assay | Delta CT (100%) with control sample \leq 5% | <3% |
| FVIII:C clotting assay | Delta CT (100%) with control sample \leq 5% | <1% |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : June 26th 2026



Quality Control Manager :

Anne-Marie VISSAC
Technical Director