



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : HEMOCLOT THROMBIN INHIBITORS

Product Code : CK002K

Lot No. : FD1045KJ

Expiration date : 2026-11-16

Components	Quantity	Lot	Expiration Date
R1: Normal Pooled Plasma	3 vials	FD1045APKJ	2026-11-16
R2 : Human Calcium Thrombin	3 vials	FD1045AQKJ	2026-11-23

Analytical Data	Specification	Result
Dabigatran standard calibration curve on STAR	Delta CT (50-500 ng / mL) : 50 ± 10 sec	55.8 sec
	$R^2 \geq 0.98$	PASS
	Controls measured in acceptance ranges	PASS
Normal Plasma	Dabigatran conc. < 10 ng / mL	PASS

Calibration Curve for information

Standard Range

Dabigatran Conc.-ng/mL	0	270	544
CT (Sec.) STAR	27.4	59.3	95.0
CT (Sec.) CS	41.3	78.9	130.9

QC:	TV* (ng/ml)	MV* (ng/ml)
C1	[82-124]	102
C2	[212-286]	238

Low Range

Dabigatran Conc.-ng/mL	0	43	110
CT (Sec.) STAR	26.4	48.3	82.2
CT (Sec.) CS	39.9	59.4	95.4

QC:	TV* (ng/ml)	MV* (ng/ml)
C1	[15-35]	25
C2	[40-80]	63

* TV= Target Value - MV= Measured Value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

20 JUN 2024

S. LECOURT

Quality Control Manager :