



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST PAI-1 Activity

Product Code : RK019A

Lot No. : FD1564

Expiration date : 2026-11-30

Components	Quantity	Lot	Expiration Date
tPA pre-coated plate	12x8 wells	FD1578	2028-01-24
(h) PAI-1 standard	3 vials	FD1597	2028-01-31
(h) PAI-1 Control I	1 vial	FD1602	2028-01-31
(h) PAI-1 Control II	1 vial	FD1604	2028-01-31
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	FD1599	2028-02-03
F-Sample diluent	2x50 ml	FD1565	2028-01-19
Wash solution	1x50 ml	FD1329	2027-12-22
Conjugate diluent	1x25 ml	FD0391	2027-08-02
TMB	1x25 ml	220106D03R	2026-11-30
Sulfuric Acid 0,45M	1x6 ml	FD1504	2028-01-12

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml \geq 1.00	1.54
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.01
Detection threshold	\leq 0.5 ng/ml	< 0.1 ng/ml

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.775	1.033	0.578	0.259	0.137

	Acceptance Range	MV
CI	[8.7 - 11.7]	10
CII	[2.1 - 3.1]	2.6

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2024/09/10

Quality Control Manager : Isabelle CORNUEJOLS
Resp. Labo CQ

