

 HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ZYMUTEST FACTOR VII

**Product Code :** RK036A

**Lot No. :** FD1874

**Expiration date :** 2026-11-20

Components	Quantity	Lot	Expiration Date
Anti (h) Factor VII pre-coated plate	12x8 wells	FD1906	2027-03-19
Factor VII standard	3 vials	FD1094	2026-11-20
Factor VII Control I	1 vial	FD1092	2027-11-15
Factor VII Control II	1 vial	FD1090	2027-11-15
Anti-(h) Factor VII HRP immunoconjugate	3 vials	FD1184	2026-12-04
Sample diluent	2x50 ml	FD0573	2027-08-30
Wash solution	1x50 ml	FD0322	2027-07-26
Conjugate diluent	1x25 ml	FD0391	2027-08-02
TMB	1x25 ml	221201D01R	2027-10-31
Sulfuric Acid 0,45M	1x6 ml	FC1736	2027-01-04

Analytical Data	Specification	Result
Reactivity	A450 for 100 % $\geq$ 1.00	1.59
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.057
Detection threshold	$\leq$ 5%	< 1%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

**24 OCT. 2024**  
**S. LECOURT**

Quality Control Manager :

**Sophie LECOURT**  
Deputy Director Quality Control

