



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ARGATROBAN CONTROL PLASMA

Product Code : SC035K

Lot No. : FD11001B

Expiration date : 2026-11-21

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD1100CE1B	2026-11-21
C2	6 vials	FD1100CF1B	2026-11-21

Analytical Data	Specification	Result
C1	0.55 - 0.80 µg/mL	0.70 µg/mL
C2	1.10 - 1.40 µg/mL	1.26 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

03.07.2024

Quality Control Manager :

Claire DUNOIS
Dir. Etudes Clin. & Val.

