




 HYPHEN BioMed  
 155 rue d'Eragny  
 95000 Neuville-sur-Oise  
 France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN NORMAL CONTROL PLASMA

**Product Code :** 223201

**Lot No. :** FD2324PC

**Expiration date :** 2026-12-07

Components	Quantity	Lot	Expiration Date
Normal Control	12 vials	FD1190DZKS	2026-12-07

Analytical Data	Specification	Result
INR	<1.3	PASS
aPTT LR Ratio	≤ 1.20	PASS
Fibrinogen (Activity)	> 2 g/L	PASS
Fibrinogen (Antigen)	> 2 g/L	PASS
Thrombin time Ratio	≤ 1.20	PASS
LA ratio	≤ 1.2	PASS
Alpha 2 Anti plasmin	≥ 80 %	PASS
Antithrombin (Activity)	> 80 %	PASS
Antithrombin (Antigen)	> 80 %	PASS
Factor II	> 70 %	PASS
Factor V	> 70 %	PASS
Factor VII	> 70 %	PASS
Factor X	> 70 %	PASS
Factor VII+ X	> 70 %	PASS
Factor VIII	> 60 %	PASS
Factor IX	> 60 %	PASS
Factor XI	> 70 %	PASS
Factor XII	> 70 %	PASS
Factor XIII	> 60 %	PASS
Free Protein S (Antigen)	> 70 %	PASS
Plasminogen	> 75 %	PASS
Protein C Chromogenic	> 80 %	PASS
Protein C (Antigen)	> 70 %	PASS
Protein C Clotting	> 60 %	PASS
Protein S (Activity)	> 50 %	PASS
von Willebrand Factor (Antigen)	> 60 %	PASS
Act. PC-r (quantitative)	< 10 % FVL	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2024-12-19

Quality Control Manager :

P/O S. LÉTOURAT  
