



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN LMWH CONTROL C3
Product Code: 223801
Lot No.: FD1325L9
Expiration date: 2026-12-29

Components	Quantity	Lot	Expiration Date
C3	12 vials	FD1325BOL9	2026-12-29

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml	0.77 IU/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: **Quality Control Approval** **Quality Control Manager:**
25 JUL. 2024
S. LECOURT

