



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST Protein C

Product Code : RK027A

Lot No. : FD1614

Expiration date : 2026-12-16

Components	Quantity	Lot	Expiration Date
Anti (h) Protein C pre-coated plate	12x8 wells	FD1846	2027-03-13
Protein C Standard	3 vials	FD1610	2027-02-08
Protein C Control I	1 vial	FD1615	2027-02-08
Protein C Control II	1 vial	FD1617	2027-02-08
Anti-(h)-Protein C HRP immunoconjugate	3 vials	FD1903	2027-03-19
Sample diluent	2x50 ml	FD0572	2027-08-30
Wash solution	1x50 ml	FD0331	2027-07-26
Conjugate diluent	1x25 ml	FC1583	2026-12-16
TMB	1x25 ml	240508D01	2028-05-31
Sulfuric Acid 0,45M	1x6 ml	FD0463	2027-08-09

Analytical Data	Specification	Result
Reactivity	A450 for 100% ≥ 1.40	2.01
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.016
Detection threshold	$\leq 5 \%$	< 1

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20	Acceptance Range	MV
OD	2.175	1.762	1.29	0.707	0.422	CI [62 - 84]	74
						CII [33 - 49]	42

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

23 OCT. 2024

S. LECOURT

