

HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ZYMUTEST vWF

**Product Code :** RK030A

**Lot No. :** FD0756

**Expiration date :** 2026-12-31

Components	Quantity	Lot	Expiration Date
Anti (h) vWF pre-coated plate	12x8 wells	FD1060	2027-11-08
vWF standard	3 vials	FD0757	2027-09-23
vWF Control I	1 vial	FD0759	2027-09-23
vWF Control II	1 vial	FD0760	2027-09-23
Anti-(h)-vWF HRP immunoconjugate	3 vials	FD1058	2027-11-11
Sample diluent	2x50 ml	FD0572	2027-08-30
Wash solution	1x50 ml	FD0073	2027-06-22
Conjugate diluent	1x25 ml	FD0390	2027-08-02
TMB	1x25 ml	221201D01	2026-12-31
Sulfuric Acid 0,45M	1x6 ml	FD0463	2027-08-09

Analytical Data	Specification	Result
Reactivity	A450 for 100 % $\geq$ 1.40	1.599
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.013
Detection threshold	$\leq$ 5%	<1%

### Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.819	1.329	0.867	0.400	0.214


	Acceptance Range	MV
CI	[ 65 - 88 ]	79
CII	[ 35 - 52 ]	43

MV = Measured value

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**

Quality Control Manager :



20 JUN 2024

S. LECOURT