



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST FACTOR X

Product Code : RK033A

Lot No. : FD0297

Expiration date : 2026-12-01

Components	Quantity	Lot	Expiration Date
Anti (h) Factor X pre-coated plate	12x8 wells	FD0397	2027-08-02
Factor X standard	3 vials	FD0333	2027-07-27
Factor X Control I	1 vial	FD0339	2027-07-27
Factor X Control II	1 vial	FD0335	2027-07-27
Anti-(h) Factor X HRP immunoconjugate	3 vials	FD0399	2027-08-05
Sample diluent	2x50 ml	FC1498	2026-12-01
Wash solution	1x50 ml	FD0322	2027-07-26
Conjugate diluent	1x25 ml	FD0391	2027-08-02
TMB	1x25 ml	221201D01	2026-12-31
Sulfuric Acid 0,45M	1x6 ml	FC2199	2027-03-30

Analytical Data	Specification	Result
Reactivity	A450 for 100 % \geq 1.40	1.607
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	$<$ 0.100	0.018
Detection threshold	\leq 5%	$<$ 1%

Calibration Curve for information

						Acceptance Range	MV	
Conc.	C	C/2	C/4	C/10	C/20	CI	[65 - 88]	78
OD	1.709	1.264	0.841	0.425	0.238	CII	[27 - 41]	31

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

07 MARS 2024

S. LECOURT

