



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF: CBA

Product Code : RK038A

Lot No. : FD0755

Expiration date : 2026-12-16

Components	Quantity	Lot	Expiration Date
Collagen pre-coated plate	12x8 wells	FD1004	2027-11-02
vWF standard	3 vials	FD0753	2027-09-23
vWF Control I	1 vial	FD0749	2027-09-23
vWF Control II	1 vial	FD0751	2027-09-23
Anti-(h)- vWF HRP immunoconjugate	3 vials	FD1062	2027-11-11
Sample diluent	2x50 ml	FD0572	2027-08-30
Wash solution	1x50 ml	FD0073	2027-06-22
Conjugate diluent	1x25 ml	FC1583	2026-12-16
TMB	1x25 ml	221201D01	2026-12-31
Sulfuric Acid 0,45M	1x6 ml	FD0463	2027-08-09

Analytical Data	Specification	Result
Reactivity	A450 for 100 % \geq 1.00	1.95
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.001
Detection threshold	\leq 5%	<1%

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.123	1.522	0.931	0.433	0.232

	Acceptance Range	MV
CI	[49 - 66]	57
CII	[24 - 36]	29

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager : *P.O. 27.06.2024*

11 JUN 2024

S. LECOURT