



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL LOW
Product Code : 223701
Lot No. : FD1622
Expiration date : 2027-01-04

Components	Quantity	Lot	Expiration Date
CI	6 vials	FD1398BQLH	2027-01-04
CII	6 vials	FD1400BRLJ	2027-01-04

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml CII : 0.35 -0.65 IU/ml	0.26 IU/mL 0.52 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **12 AOUT 2024**

Quality Control Manager :

Claire DUNOIS
Dir. Etudes Clin. & Val.