



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN UFH CONTROL C1  
**Product Code :** 224101  
**Lot No. :** FD1434LM  
**Expiration date :** 2027-01-09

Components	Quantity	Lot	Expiration Date
C1	12 vials	FD1434BXML	2027-01-09

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml	0.18 IU/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 12 AOUT 2024

Quality Control Manager :

Claire DUNOIS  
Dir. Etudes Clin. & Val.