



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

CERTIFICATE OF ANALYSIS	
<b>Product Description :</b>	<b>BIOPHEN HEPARIN CALIBRATOR</b>
<b>Product Code :</b>	<b>222001</b>
<b>Lot No. :</b>	<b>FD1745MQ</b>
<b>Expiration date :</b>	<b>2027-02-27</b>

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FD1745BJMQ	2027-02-27
CAL 2	4 vials	FD1745BKMQ	2027-02-27
CAL 3	4 vials	FD1745BLMQ	2027-02-27
CAL 4	4 vials	FD1745BMMQ	2027-02-27
CAL 5	4 vials	FD1745BNMQ	2027-02-27

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 0.05$ IU/ml CAL 2 : 0.30 -0.50 IU/ml CAL 3 : 0.70 -0.90 IU/ml CAL 4 : 1.10 -1.30 IU/ml CAL 5 : 1.45 -1.75 IU/ml	0 IU/mL 0.40 IU/mL 0.81IU/mL 1.20 IU/mL 1.59 IU/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

**Quality Control Approval**

Date of Approval :

01 OCT. 2024  
S. LECOURT

Quality Control Manager :