



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS
Product Description : ZYMUTEST HIA IgG,A,M
Product Code : RK040D
Lot No. : FD1277
Expiration date : 2027-02-12

Components	Quantity	Lot	Expiration Date
Heparin pre-coated plate	12x8 wells	FD1805	2027-03-05
HIA Ig G,A,M Positive Control	3 vials	FD1853	2027-03-12
Negative Control	3 vials	FD1266	2027-12-10
Anti (h) IgG - Anti IgA - Anti IgM HRP immunoconjugate	3 vials	FD1852	2028-03-09
HIA Sample diluent	2x50 ml	FD1631	2027-02-12
Wash solution	1x50 ml	FD1330	2027-12-22
Cell Lysate	3 vials	FD1803	2027-03-08
Conjugate diluent	1x25 ml	FD1442	2028-01-03
TMB	1x25 ml	240508D01	2028-05-31
Sulfuric Acid 0,45M	1x6 ml	FD1506	2028-01-12

Analytical Data	Specification	Result
Reactivity	A450 for Positive Control ≥ 1.00 A450 for Negative Control ≤ 0.30	2.169 0.176
Pathological Plasmas	A450 ≥ 0.500	PASS
Blank	< 0.100	PASS
Normal Plasma	M + 2SD ≤ 0.50	0.312

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

15 OCT. 2024

Quality Control Manager :

S. LECOURT